

## TENDER FOR "SUPPLY, INSTALLATION AND AMC OF FUEL GAS ANALYZER for HEAT & BOILER CHEMISTRY LAB" AT INDIAN MARITIME UNIVERSITY, MUMBAI PORT CAMPUS

TENDER No: IMU-MPC/PUR/2020-21/H & C Lab./12

Issue Date: 20th April, 2021

## **CORRIGENDUM - II**

Following corrigendum is issued with respect to the "SUPPLY, INSTALLATION AND AMC OF FUEL GAS ANALYZER for HEAT & BOILER CHEMISTRY LABORATORY" tender:

## 1. Page No.7, para No.11 (11.1) table amended as follows:

Sr. No.	Name of Equipment	Technical Details	Parameter	Range	Resolution	Qty
1	Exhaust Gas	1. Instrument can measure -CO,	O <sub>2</sub>	0-25%	0.1%	01
	Analyzer- portable type	O <sub>2</sub> ,SO <sub>x</sub> ,NO <sub>x</sub> ,CO <sub>2</sub> cal culated valued,	CO measurement (H2 compensated)	0-5000ppm	1 ppm	
		2. Portable type instrument	NO measurement	0-3000ppm	1 ppm	
		3. Electronic reading which can be	NO2 measurement	0-500 ppm	1 ppm	
		imported to Computer for print	SO2 measurement	0-5000ppm	1ppm	
		out.	CO2 measurement (Calculated from	0-25%	0.1%	
			O2) Temp Measurement	0-1000 degree C	1 degee c	